



**microINR portable coagulometer**

A system for measurement P—Prothrombin time (INR)  
manufactured by iLine Microsystems S.L.

**Report from the evaluation SKUP/2015/109**

*organised by SKUP at the request of iLine Microsystems S.L.*

SKUP in Denmark, Dept. KBA, Nordsjællands Hospital, 3400 Hillerød, Phone +45 4829 4176, [www.skup.nu](http://www.skup.nu)  
SKUP in Norway, Noklus, Box 6165, NO-5892 Bergen, Phone +47 55979502, [www.skup.nu](http://www.skup.nu)  
SKUP in Sweden, Equalis, Box 977, SE-751 09 Uppsala, Phone: +46 18 490 31 44, [www.skup.nu](http://www.skup.nu)

**SKUP secretariat**

Grete Monsen  
+47 55 97 95 02  
grete.monsen@noklus.no

**SKUP in Denmark**

Esther Jensen  
Department of Clinical Biochemistry  
Nordsjællands Hospital  
Dyrehavevej 29, indgang 16A  
DK-3400 Hillerød  
+45 48 29 41 76  
+45 24 82 28 36 (cell phone)  
esther.jensen@regionh.dk

**SKUP in Norway**

Grete Monsen  
Marianne Risa  
Anne Christin Breivik  
Sverre Sandberg  
Noklus  
Boks 6165  
NO-5892 Bergen  
+47 55 97 95 02  
grete.monsen@noklus.no  
marianne.risa@noklus.no  
anne.breivik@noklus.no  
sverre.sandberg@noklus.no

**SKUP in Sweden**

Elisabet Eriksson Boija  
Gunnar Nordin  
Equalis AB  
Box 977  
SE-751 09 Uppsala  
+46 18 490 31 44  
elisabet.eriksson.boija@equalis.se  
gunnar.nordin@equalis.se

**www.SKUP.nu**

*SKUP would like to acknowledge with thanks those who contributed to the practical work with this evaluation including Caroline Arnesen, Kari Bratberg, Lillian Ottem Dahl, Siv Furunes, Margrete Kviseth Gulbrandsen, Guri Andersen Gulstad, Kirsten Halvorsen, Linn Silje Hansen, Hilde Hegseth, Åse Jekthammer, Elise Nyberg, Line Nygård, Hege Elin Skogan, Kristin Sollihaug and Tonje Wold.*

The report was written by SKUP, autumn 2015. Main author was Anne Christin Breivik, SKUP in Norway.

---

In order to use the SKUP name in marketing, it has to be referred to [www.skup.nu](http://www.skup.nu) and to the report code in question; SKUP/2015/109. For more details, see attachment 1.

Table of contents

**1. SUMMARY.....4**

**2. ABBREVIATIONS AND ACRONYMS .....5**

**3. INTRODUCTION.....6**

**3.1. BACKGROUND FOR THE EVALUATION.....6**

**3.2. THE AIM OF THE EVALUATION .....6**

**3.3. THE SKUP MODEL .....6**

**4. QUALITY GOALS .....7**

**4.1. ANALYTICAL QUALITY .....7**

**4.2. USER-FRIENDLINESS.....8**

**4.3. PRINCIPLES FOR THE ASSESSMENTS .....8**

**4.4. SKUP’S QUALITY GOALS IN THIS EVALUATION .....9**

**5. MATERIALS AND METHODS.....10**

**5.1. DEFINITION OF THE MEASURAND.....10**

**5.2. THE EVALUATED MEASUREMENT SYSTEM MICROINR .....10**

**5.3. THE SELECTED COMPARISON METHOD.....12**

**5.4. THE EVALUATION.....13**

**6. RESULTS AND DISCUSSION.....16**

**6.1. NUMBER OF SAMPLES .....16**

**6.2. ANALYTICAL QUALITY OF THE SELECTED COMPARISON METHOD.....17**

**6.3. ANALYTICAL QUALITY OF THE MICROINR SYSTEM UNDER OPTIMAL CONDITIONS .....19**

**6.4. ANALYTICAL QUALITY OF THE MICROINR SYSTEM ACHIEVED BY INTENDED USERS .....22**

**6.5. EVALUATION OF USER-FRIENDLINESS .....26**

**7. REFERENCES .....31**

**ATTACHMENTS**

1. The organisation of SKUP
2. Facts about the microINR system
3. Information about manufacturer, retailers and marketing
4. Product specifications for this evaluation, microINR system
5. Statistical expressions and calculations
6. Raw data PT (INR), results from the comparison method
7. Raw data PT (INR), microINR internal analytical quality control results, optimal conditions
8. Raw data PT (INR), microINR results, optimal conditions
9. Raw data PT (INR), microINR internal analytical quality control results, intended users
10. Raw data PT (INR), microINR results, intended users
11. “SKUP-info”. Summary for primary health care (in Norwegian)
12. List of previous SKUP evaluations
13. Comments from iLine Microsystems S.L.

Attachments with raw data are included only in the copy to iLine Microsystems S.L.

## 1. Summary

### Background

The microINR portable coagulometer is an in vitro diagnostic device for the quantitative measurement of prothrombin time, PT (INR). The system is intended for professional use, self-testing and self-management of patients on oral vitamin K antagonist therapy. The sample material is fresh capillary whole blood. The system is produced by iLine Microsystems S.L. and was launched into the Scandinavian market December 2012. The SKUP evaluation was carried out from March to August 2015 at the request of iLine Microsystems S.L. in Spain.

### The aim of the evaluation

The aim of the evaluation was to determine the analytical quality and user-friendliness of the microINR system, both when used under optimal conditions in a hospital laboratory and by intended users in primary health care. The analytical results were assessed according to the quality goals set for the evaluation.

### Materials and methods

In a hospital laboratory, capillary samples from 98 patients were measured with the microINR system. In two primary health care centres, capillary samples from 40 and 48 patients respectively were measured with the microINR system. Venous samples from the same patients were analysed on the comparison method implemented on STA-R Evolution. The quality goal was a repeatability (CV)  $\leq 5,0\%$  and for accuracy that  $\geq 95\%$  of the results should be within  $\pm 20\%$  from the results of the comparison method. The quality goal for the user-friendliness was a total rating of “satisfactory” including an incident of tests wasted due to technical errors  $\leq 2\%$ . The microINR system was tested in stationary mode, i.e. not moved during sample application or measurement.

### Results

For PT (INR) results  $< 2,5$  the repeatability was estimated to be just below  $5,0\%$  at all evaluation sites. For PT (INR) results  $\geq 2,5$  the repeatability was  $6,0\%$  under optimal conditions and between  $6,1\%$  and  $6,3\%$  achieved by the intended users. At PT (INR) level  $< 2,5$  there was a statistically significant bias of  $+0,06$  INR under optimal conditions. For PT (INR) level  $\geq 2,5$  there was no statistically significant bias. Under optimal conditions,  $97\%$  of the results were within the quality goal for accuracy, and when handled by the intended users  $95\%$  of the samples were within the quality goal for accuracy. The user-friendliness was rated as intermediate and the incident of tests wasted due to technical errors was  $1,6\%$ .

### Conclusion

The quality goal for repeatability at PT (INR) level  $< 2,5$  was most likely fulfilled. At PT (INR) level  $\geq 2,5$  the quality goal for the repeatability was not fulfilled. The quality goal for accuracy was fulfilled. The quality goal for user-friendliness was not fulfilled.

### Comments from iLine Microsystems S.L.

A letter with comment from iLine Microsystems S.L. is attached to the report.

## **2. Abbreviations and Acronyms**

BLS	Biomedical Laboratory Scientist
CI	Confidence Interval
C-NPU	The committee on Nomenclature, Properties and Units
CV	Coefficient of Variation
DEKS	Danish Institute of External Quality Assurance for Laboratories in Health Care
EQA	External Quality Assessment
Equalis	External quality assurance in laboratory medicine in Sweden
NKK	Norwegian Clinical Chemistry EQA Program
Noklus	Norwegian Quality Improvement of Primary Care Laboratories
NS_EN ISO/ IEC	Norsk Standard_Europeisk Norm International Organization for Standardization/International Electrotechnical Commission
PHCC	Primary health care centre
PT (INR)	Prothrombin Time International Normalized Ratio
RBT	Rabbit Brain Thromboplastin
SKUP	Scandinavian evaluation of laboratory equipment for primary health care
WHO	World Health Organization

### 3. Introduction

#### 3.1. Background for the evaluation

The microINR portable coagulometer is produced by iLine Microsystems S.L. and was launched into the Scandinavian market December 2012. iLine Microsystems S.L. is the requesting company in this evaluation. A first SKUP evaluation of the microINR system was organized during spring 2014. This evaluation was stopped halfway on the request from the producer, due to underperformance of an R&D Software version provided for the evaluation.

#### 3.2. The aim of the evaluation

The aim of the evaluation was to determine the analytical quality and user-friendliness of the microINR system, both when used under optimal conditions in a hospital laboratory and by intended users in primary health care.

The evaluation includes:

- Examination of the analytical quality (precision and accuracy) under optimal conditions
- Examination of the analytical quality (precision and accuracy) in the hands of intended users
- Evaluation of the user-friendliness of the microINR system and manual

#### 3.3. The SKUP model

SKUP evaluations for quantitative methods are based upon the fundamental guidelines in a book concerning evaluations of laboratory equipment in primary health care [1]. A complete SKUP evaluation consists of two parts. One part of the evaluation is carried out under optimal conditions in a hospital laboratory. This part documents the quality of the system under conditions as favourable as possible for achieving good analytical quality. The other part of the evaluation is carried out by intended users in at least two primary health care centres (PHCCs). This part documents the quality of the system under real-life conditions.

The evaluation under optimal conditions includes:

- Repeatability with 100 patient samples
- Comparison with an established hospital laboratory method

The evaluation performed by intended users includes:

- Repeatability with 40 patient samples at each of the primary health care centres
- Comparison with an established hospital laboratory method
- Evaluation of user-friendliness

If possible, SKUP evaluations are carried out using three lot numbers of test chips from separate and time-spread productions.

## 4. Quality goals

### 4.1. Analytical quality

At the present, there are no generally recognised analytical quality goals for the determination of prothrombin time International Normalized Ratio (PT (INR)), and no international (Gold) standard for evaluation of Point of Care test instruments for PT (INR) in primary health care.

The International Organization for Standardization (ISO) 17593 standard [2] gives requirements for monitoring systems for self-testing of oral anticoagulant therapy. In SKUP's opinion, the quality goals for accuracy in the ISO 17593 standard ( $\pm 30\%$  for 90% of the results in the therapeutic range 2 – 4,5 PT (INR) ) is too tolerant. Furthermore, there is no performance criterion for imprecision in the standard.

Setting quality goals based on biological variation is an acknowledged method [3,4]. It is recommended that analytical imprecision (repeatability, coefficient of variation  $CV_a$ ) should be less than, or equal to, half the intra-individual biological variation. Ricos *et al.* [5] state the biological variation for prothrombin time as  $CV_{bw}$  4% (intra-individual biological variation) and  $CV_{bb}$  6,8% (inter-individual biological variation). According to Kjeldsen *et al.* [6], the “in-treatment within-subject biological variation” of PT (INR) is 10,1%. For systems used for monitoring, the analytical performance should aim at low imprecision compared to the within-subject biological variation. In principle, quality goals based on biological variation do not take into account clinical requirements.

A committee appointed by the National Ministry of Health in Denmark has specified the requirements of analytical quality for PT (INR) [7,8] to bias  $\leq 6\%$  and imprecision  $\leq 5\%$  for instruments used in primary health care, and bias  $\leq 3\%$  and imprecision  $\leq 3\%$  for hospital instruments. There is no separate goal for the total error in the Danish specifications; however, estimated CV for the matrix effect is defined and an allowable deviation is given in the control system.

Based on the given data on biological variation for prothrombin time, and the fact that PT (INR) devices used in primary health care are designed for *monitoring* prothrombin time, SKUP recommends that these instruments should achieve a repeatability ( $CV$ )  $\leq 5,0\%$ . SKUP has not taken out a separate goal for bias, but a figure of 5% was used to calculate a quality goal for allowable deviation according to the model below.

In method evaluations and comparisons, the imprecision of the comparison method has to be taken into account. SKUP allows an imprecision ( $CV$ ) of the comparison method up to 3%. In addition, inter-laboratory variation should be taken into the calculation of the allowable deviation, which SKUP has estimated to 3%.

When comparing two prothrombin time methods, especially when the methods represent two different measuring principles, certain sample specific errors can be assumed. SKUP has chosen to include a variation of 5% in the error model for calculation of allowable deviation.

$$\begin{aligned} \text{Allowable deviation} &= |\pm\text{bias}| + 1,65 \times \sqrt{CV_{\text{test method}}^2 + CV_{\text{comparison method}}^2 + CV_{\text{betweenlab}}^2 + CV_{\text{matrix}}^2} \\ &= (5 + 1,65 \times \sqrt{25 + 9 + 9 + 25}) = (5 + 13,6) \approx \pm 19\% \end{aligned}$$

## 4.2. User-friendliness

The evaluation of user-friendliness is carried out by asking the evaluating persons to fill in a questionnaire divided into four subareas, see section 6.5.

### *Technical errors*

SKUP recommends that the percentage of “tests wasted” caused by technical errors should not exceed 2%.

## 4.3. Principles for the assessments

To qualify for an overall good assessment in a SKUP evaluation, the measuring system must show satisfactory analytical quality as well as satisfactory user-friendliness.

### 4.3.1. Assessment of the analytical quality

The analytical results are assessed according to the quality goals set for the evaluation.

### *Precision*

The decision whether the achieved CV fulfills the quality goal or not is made on a 5% significance level. The distinction between the ratings, and the assessment of precision according to the quality goal, are shown in table 1.

**Table 1.** The rating of precision

Distinction between the ratings	Assessment according to the quality goal
The CV is lower than the quality goal (statistically significant)	The quality goal is fulfilled
The CV is lower than the quality goal (not statistically significant)	Most likely the quality goal is fulfilled
The CV is higher than the quality goal (not statistically significant)	Most likely the quality goal is not fulfilled
The CV is higher than the quality goal (statistically significant)	The quality goal is not fulfilled

### *Trueness*

The confidence interval (CI) of the measured bias is used for deciding if a difference between the two methods is statistically significant (two-tailed test, 5% significance level). The term trueness is related to the results achieved under optimal conditions. Proven systematic deviation of the results achieved by intended users will be discussed in relation to the bias found under optimal conditions.



*Accuracy*

The accuracy is illustrated in a difference plot with limits for the allowable deviation according to the quality goal. The fraction of results within the limits is counted. The accuracy is assessed as either fulfilling the quality goal or not fulfilling the quality goal.

*Bias with three lots of test chips*

Separate lot calculations are not performed. The results achieved with the three lots are included in the assessment of accuracy in the difference plots. If distinct differences between the lots appear, this will be pointed out and discussed.

**4.3.2. Assessment of the user-friendliness**

The user-friendliness is assessed according to the answers and comments given in the questionnaire (see section 6.5.). For each question, the evaluator must choose between three given ratings. The responses from the evaluators are reviewed and summed up. To achieve the overall rating “satisfactory”, the tested equipment must reach the total rating of “satisfactory” in all four subareas of characteristics described in section 6.5.

*Technical errors*

The evaluating person registers error codes and technical errors during the evaluation. The fraction of technical errors is calculated and taken into account in connection with the assessment of the user-friendliness.

**4.4. SKUP’s quality goals in this evaluation**

As agreed upon when working on the protocol, the results from the evaluation of the microINR system are assessed against the following quality goals:

Repeatability (CV) .....	≤5,0%
Allowable deviation in the individual result from the comparison method result.....	<+20%
Required percentage of individual results within the allowable deviation* .....	≥95%
Fraction of technical errors .....	≤2%
User-friendliness, overall rating.....	Satisfactory

\*If more than 1% of the results deviate more than ±25%, this will be pointed out and discussed

## 5. Materials and methods

### 5.1. Definition of the measurand

The Committee on Nomenclature, Properties and Units (C-NPU) describes clinical laboratory tests in a database [9]. In the NPU-database the specifications for the measurand in this evaluation are as shown in table 2.

**Table 2.** NPU-specifications

NPU code	Name of test according to NPU	Unit
NPU01685	P—Coagulation, tissue factor-induced; relative time(actual/normal; INR; IRP 67/40; proc.)	–
NPU21717	P—Coagulation, tissue factor-induced; rel.time(actual/norm; INR; IRP 67/40; II+V+VII+X)	–

The analytical test according to NPU01685 refers to measurements performed with the Owren method. The test is mainly determined by the concentration of the Vitamin K dependent coagulation factors II, VII and X. The analytical test according to NPU21717 refers to measurements performed with the Quick method. The test is mainly determined by the concentration of the Vitamin K dependent coagulation factors, in addition to fibrinogen (factor I) and factor V.

Even if the tests according to NPU01685 and NPU21717 are not measuring exactly the same plasma components, the test results are used as if they did. In this report, the comparison method is an Owren method while the evaluated method, the microINR system, is a Quick method. The term “PT (INR)” will be used for the measurand in this report. As the measurement result is a ratio of the actual coagulation time divided with the normal coagulation time (INR), there is no unit.

### 5.2. The evaluated measurement system microINR

The information in this section derives from the company information material.

The microINR system is intended for the quantitative determination of PT (INR). The microINR meter (figure 1) and the disposable analytical microINR test chips compose the microINR portable coagulometer. The product is intended for professional use, self-testing and self-management of patients on oral vitamin K antagonist therapy.



**Figure 1.** microINR

The sample material on the microINR system is fresh capillary whole blood. The Rightest GD500 Lancing device is supplied for sampling. This device is for repeated use on a single person only and it was not used in this evaluation.

The microINR chips have two channels; one for sample measuring and one for control. Each micro capillary channel consists of a reaction chamber, where the reagent is placed, and a micro capillary, where PT (INR) is determined. The measuring channel contains dried reagent of human recombinant thromboplastin in a phospholipidic matrix and stabilizers, and so does the control channel, but in addition it contains human clotting factors.

Blood is inserted into the chip from the entry channel; it is divided into the two channels and mixed with the reagents, which activates the coagulation cascade. When blood clotting has occurred, PT (INR) calculation is performed from the monitored curves.

The calibration of the test is done automatically as the International Sensitivity Index (ISI) and Mean Normal Prothrombin Time (MNPT) value for the lot of the microINR chips is coded in the data matrix printed on each chip.

The manufacturer iLine Microsystems S.L. produces the control material microINR Easy Control. The material is lyophilised human abnormal plasma with buffer, stabilizers and preservatives.

According to the manual of the microINR system, there are two modes for sample application:

A: Approach finger to the meter

B: Approach the meter to the pricking area

This evaluation was performed using mode A, as instructed by the producer and supplier.

For technical details about the microINR system, see table 3. For more technical information about the microINR system, name of the manufacturer and the suppliers in the Scandinavian countries, see attachment 2 and 3. For product specifications in this evaluation, see attachment 4.

**Table 3.** Technical details from the manufacturer

<b>Technical details for the microINR system</b>	
Sample material	Fresh capillary blood
Sample volume	3 $\mu$ L
Measuring time	
Measuring range	0,8 – 8,0 INR
Haematocrit	25% – 55%
Storage capacity	199 results
Electrical power supply	Rechargeable battery

### 5.3. The selected comparison method

A selected comparison method is a fully specified method, which in the absence of a Reference method, serves as a common basis for the comparison of the evaluated method.

#### 5.3.1. The selected comparison method in this evaluation

The selected comparison method in this evaluation is the routine method for PT (INR) at the Department of Medical Biochemistry at St. Olavs Hospital in Trondheim, Norway, hereafter called “the comparison method”.

The method is accredited after NS\_EN ISO/IEC 17025 (Norsk Standard\_Europeisk Norm International Organization for Standardization /International Electrotechnical Commission, 2005).

<i>Instrument:</i>	STA-R Evolution, STAGO, two identical instruments were in use	
<i>Reagent:</i>	STA-SPA+, Diagnostica STAGO	
<i>Principle:</i>	Owren’s method, rabbit brain thromboplastin and adsorbed bovine plasma	
<i>Traceability:</i>	World Health Organization’s (WHO’s) manual tilt tube technique and the reference thromboplastin WHO IRP 67/40, through Rabbit Brain Thromboplastin (RBT/90) [9-11]	
<i>Calibrators:</i>	Three point’s calibration with PT (INR)-calibrators from Equalis (External quality assurance in laboratory medicine in Sweden)	
<i>Reference range</i>	0,9 – 1,2 INR	
<i>Therapeutic range</i>	venous indication	2,0 – 3,0 INR
	arterial indication	2,5 – 3,5 INR

#### *Internal analytical quality control*

Internal analytical quality control samples, two levels (STA-Scandinorm PT (INR) and STA-Scandipath PT (INR), STAGO) were measured each evaluation day on the comparison method instruments. The reproducibility of the comparison method as achieved with the quality control material was calculated.

#### *External analytical quality control*

The hospital laboratory participates in Noklus/NKK (Norwegian Quality Improvement of Primary Care Laboratories/Norwegian Clinical Chemistry EQA Program) external quality assessment (EQA) scheme for PT (INR) with two levels four rounds per year. The materials are freshly frozen pooled citrate plasma from Norwegian donors. The assigned value for PT (INR) is based on consensus values from 67–69 participants using PT (INR)-calibrators from Equalis.

#### 5.3.2. Verification of the analytical quality of the comparison method

##### *Precision*

The repeatability of the comparison method was calculated from duplicate measurements of venous citrate samples from patients in stable ( $\geq 4$  weeks) vitamin K antagonist treatment. The requested repeatability of the laboratory is 2,4% at a PT (INR) level of approximately 3,0.

### *Trueness*

- PT (INR) calibrators from Equalis were analysed as samples on the comparison method at the start and in the end of the evaluation. The calibrator material is a pool of citrated anti-coagulated freeze-dried plasma of human origin (Swedish donors). The certified values are traceable to an internationally agreed reference measurement procedure (WHO's manual tilt tube technique) and the reference thromboplastin WHO IRP 67/40, through RBT/90 [10-12]. The procedures used to assign values are described in several publications and documents [13,14].
- PT (INR) calibrators from the Danish Institute for External Quality Assurance for Hospital Laboratories (DEKS) were used to get a link to the Danish PT (INR) level. The calibration materials from DEKS are freshly frozen pooled citrate-plasmas, which serve as national reference plasmas in Denmark. The DEKS calibration is a three point's calibration with a normal, therapeutic and high PT (INR). The assigned values come from three Nordic expert laboratories. The calibrators were analysed as samples on the comparison method at the start, in the middle and in the end of the evaluation.

The calibrating systems from Equalis and DEKS are different with respect to the production of the materials as well as to the way the PT (INR)-values are set.

- At different occasions during the evaluation period, PT (INR) controls from Noklus were analysed on the comparison method.

## **5.4. The evaluation**

### **5.4.1. Planning of the evaluation**

#### *Inquiry about an evaluation*

iLine Microsystems S.L. via marketing manager Clara Grijelmo, applied to SKUP in November 2014 for an evaluation of the microINR system.

#### *Protocol, arrangements and contract*

In March 2015, the protocol for the evaluation was approved, and iLine Microsystems S.L and SKUP signed a contract for the evaluation. The Department of Medical Biochemistry in St. Olav Hospital, Trondheim, agreed to do the practical work with the evaluation under optimal conditions. Two PHCCs, Persaunet Legesenter and Hallset Legesenter from Sør-Trøndelag county, agreed to represent the intended users in this evaluation.

#### *Training*

Javier Alvariño from iLine Microsystems S.L, demonstrated the microINR system in the hospital laboratory. Hege Anette Martinsen (Orion Diagnostica) demonstrated the microINR system at PHCC1 and Britt S. Fredriksen (Orion Diagnostica) demonstrated the microINR system at PHCC2. All the evaluators were instructed to use mode A (see chapter 5.2) for sample application. The training reflected the training usually given to the intended users. The requesting company and the supplier were not allowed to contact or supervise the evaluators during the evaluation period.

#### **5.4.2. Evaluation sites**

The practical work was carried out during 17 weeks at the hospital laboratory and six weeks at the PHCCs, ending in August 2015.

The laboratory at the St. Olav university hospital has approximately 100 employees.

PHCC1 has three physicians, three health secretaries and one BLS. They use venous blood samples in their routine method for measurements of PT (INR). PHCC2 has five physicians, two health secretaries and one medical secretary. They use capillary blood samples in their routine method for measurements of PT (INR).

#### **5.4.3. The evaluation procedure under optimal conditions**

##### *Internal analytical quality control*

Internal analytical quality control for the microINR system (microINR Easy Control), one level, was measured each evaluation day.

##### *Recruitment of patients*

Patients who participated in this study were patients coming to the outpatient clinic for routine PT (INR) monitoring. Blood samples were collected from patients who had been stable on vitamin K antagonist treatment for a minimum of 4 weeks. Patients with known antiphospholipid syndrome (APS) were not included. Participation was voluntarily and verbal consent was considered sufficient based on national regulations.

##### *Handling of the samples and measurements*

All samples for measurement with the microINR system were capillary samples. The skin was disinfected with alcohol pads and the area dried completely before finger pricking. The samples were measured in duplicate using two skin-pricks from two separate fingers. Disposable lancing devices with depth settings 1,8 mm were used. In this evaluation, the second drop of capillary blood was measured after wiping off the first with a clean dry tissue/gauze. The sample was applied to the chip by approaching the finger to the meter; i.e. the meter was not moved. Three lot numbers of test chips were used in the evaluation. If error codes occurred the test was repeated, if possible, until a result was obtained.

Samples for the comparison method were obtained from venous puncture and collected into vacutainer tubes with 3,2% sodium citrate. The citrate samples were taken immediately before testing of the capillary samples on the microINR system. The tubes were inverted 8–10 times to ensure thorough mixing of the blood with the sodium citrate, and then underwent centrifugation for 10 minutes at 2200 g within two hours from sampling. Citrated fresh plasma was used for duplicate measurements of PT (INR) on the comparison method.

#### **5.4.4. The evaluation procedure for intended users**

##### *Internal analytical quality control*

Internal analytical quality control for the microINR system (microINR Easy Control), one level, was measured each evaluation day.

##### *Recruitment of patients*

Patients who participated in this study were patients at the PHCCs coming for routine PT (INR) monitoring. Blood samples were collected from patients who had been stable on vitamin K antagonist treatment for a minimum of 4 weeks. Patients with known antiphospholipid syndrome

(APS) were not included. Participation was voluntarily and verbal consent was considered sufficient based on national regulations.

*Handling of the samples and measurements*

All samples for measurements with the microINR system were capillary samples. The skin was disinfected with alcohol pads and the area dried completely before finger pricking. The samples were measured in duplicate using two skin-pricks from two separate fingers. Disposable lancing devices with depth settings 1,8 mm were used. In this evaluation, the second drop of capillary blood was measured after wiping off the first with a clean dry tissue/gauze. The sample was applied to the chip by approaching the finger to the meter; i.e. the meter was not moved. Three lot numbers of test chips were used in the evaluation. If error codes occurred the test was repeated, if possible, until a result was obtained.

Samples for the comparison method were obtained from venous puncture and collected into vacutainer tubes with 3,2% sodium citrate. The citrate samples were taken immediately before testing of the capillary samples on the microINR system. The tubes were inverted 8–10 times to ensure thorough mixing of the blood with the sodium citrate. The sample tubes were transported to the Department of Medical Biochemistry according to normal routine procedures at the PHCCs. The citrate plasma was analysed in duplicate for PT (INR) on the comparison method within 48 hours after sampling.

## 6. Results and discussion

Statistical expressions and calculations used by SKUP are shown in attachment 5.

### 6.1. Number of samples

Scheduled number of samples in this evaluation were 100 patient samples in duplicate under optimal conditions and 80 patient samples in duplicate measured by intended users. The hospital recruited 98 patients (SKUP ID 1-100). In the evaluation performed by intended users, PHCC1 and PHCC2 recruited 40 and 48 patients respectively (SKUP ID 111-150 and SKUP ID 201-248).

As a total, the results were spread over a wide range, but still there were few low and high results. Most of the results were within the interval 2,0 – 3,5 INR, consequently the results achieved in the hospital laboratory were divided into two, instead of three PT (INR) levels. This also provides an easier comparison with the results achieved in the two PHCCs, where the results usually are divided in only two levels because of the lower number of results.

An account of the number of samples not included in the calculations, is given below.

#### *Missing results*

- ID 112 (E18) and ID 113 (E05); only single measurements from microINR due to error codes. The single values were included in the calculation of bias and the assessment of accuracy.
- ID 122 (E03/E05), ID 125 (E03/E18/E17) and ID 149 (E05/E18); both results missing from microINR due to error codes.
- ID 46 and ID 47; the ID numbers were not used.
- ID 146, ID 148 and ID 150, only single measurements on the comparison method. The single values from the comparison method were still included in the calculations of bias and in the assessment of accuracy.
- ID 211-217, no results from the comparison method because the evaluator placed the venous samples in the fridge. These results were not included in the calculation of bias and the assessment of accuracy, but the results from microINR were included in the calculation of repeatability.
- The hospital laboratory did not analyse the internal analytical quality control at microINR eight of the days of the evaluation. In the PHCCs internal analytical quality control result from one day was missing. The results from the patient samples these days were still included in all the calculations.

#### *Omitted results*

- ID 119; the results from microINR were not included in the calculation of repeatability due to the use of two lot numbers, i.e. not identical conditions. The venous sample was not analysed on the comparison method within 48 hours as described in the evaluation procedure. The results from this patient were not included in any calculations.
- ID 121, ID 136 and ID 148; the second measurements from microINR were omitted from all the calculations due to deviation from the sampling procedure. The first measurements were included in the calculation of bias and the assessment of accuracy.



### *Excluded results*

Statistical outliers in SKUP evaluations are detected by criterion promoted by Burnett [15]. - ID 9, ID 24 and ID 64; the results from the comparison method were classified as outliers according to Burnett's model in the calculation of repeatability of the comparison method. The comparison method had good precision, and the statistical outliers were mostly a consequence of the low CV rather than actual differences between the duplicate measurements. The results were not included in the calculation of bias and the assessment of accuracy, but the results from microINR were included in the calculation of repeatability.

### *Failed measurements*

Under optimal conditions, microINR reported the following error messages:

- 3 x E01; Meter cannot read the data matrix
- 1 x E03; Time out of the sample application countdown of 80 seconds.
- 1 x E05; Insufficient sample volume or not properly applied
- 2 x E17; Chip reading failure during testing time

Handled by intended users, microINR reported the following error messages:

- 2 x E03; Time out of the sample application countdown of 80 seconds
- 3 x E05; Insufficient sample volume or not properly applied
- 3 x E17; Chip reading failure during testing time
- 4 x E18; Sample mishandling or haematocrit out of range

Most of the error codes were related to the handling of the sample (11/372) = 3,0% errors. Eight error messages, E01 and E17, were interpreted as «technical errors», and six of these led to wasted microINR chips. The fraction of test wasted due to technical errors was estimated to:  $(6/372) \times 100 = 1,6\%$ .

The SKUP recommendation of an incident of test wasted due to technical errors  $\leq 2,0\%$  was achieved.

## **6.2. Analytical quality of the selected comparison method**

### **6.2.1. Internal analytical quality control**

Internal analytical quality control samples in two levels (STA-Scandinorm PT (INR) and STA-Scandipath PT (INR), STAGO) were measured each evaluation day on the comparison method. All results were within the allowable control limits (data not shown). After a change of reagent lot the Scandipath results became approximately 0,1 INR higher than before the change. This shift was not observed in the results of the patient samples in this evaluation. The results of Scandinorm (low level) were not affected. The reproducibility (CV) achieved with the internal analytical quality control samples were 1,5% for level 1 and 1,8–2,5% for level 2. Two lots of Scandipath were used in the evaluation.

### **6.2.2. The precision of the comparison method**

Duplicate measurements on the comparison method were performed on each of the venous citrate patient samples. The results were checked to meet the imposed condition for using formula 1 in attachment 5. There were no systematic differences pointed out between the paired measurements at the two levels (data not shown).

The precision is presented as the repeatability. The CV with a 90% CI is shown in table 4. The results are sorted and divided into two concentration levels according to the mean of the results of the comparison method. Raw data is attached for the requesting company only, attachment 6.

**Table 4.** Repeatability, PT (INR) venous citrate samples, the comparison method

PT (INR) level Comparison method	n	Excluded results	Mean value (interval), PT (INR)	CV (90% CI) %
<2,5	38	1*	2,1 (1,05 – 2,49)	1,1 (0,9 – 1,3)
≥2,5	57	2*	3,0 (2,52 – 4,67)	1,0 (0,8 – 1,2)

\* The given numbers of results (n) are counted before the exclusion of results. Mean and CV are calculated after the exclusion of results. ID 9, ID 24 and ID 64 are statistical outliers according to Burnett's model [15] and therefore excluded. An account of the number of samples is given in section 6.1.

### Discussion

The CV for the comparison method was approximately 1,0%. This is below the requested CV (2,4%) of the comparison method at a PT (INR) level of approximately 3,0.

#### 6.2.3. The trueness of the comparison method

To demonstrate the trueness of the comparison method, the calibrators from Equalis were analysed at the start and in the end of the evaluation. The calibrators from DEKS were analysed at three different occasions; at the start, in the middle and at the end of the evaluation. The results achieved with the Equalis calibrators are shown in table 5. The results achieved with DEKS calibrators are shown in table 6.

**Table 5.** Equalis PT (INR) calibrators measured on the comparison method

Material	Certified value, PT (INR) (uncertainty)	Date	n	Mean value, PT (INR) instrument 1	Mean value, PT (INR) instrument 2
Equalis INR calibrator Low	<b>1,05</b> (0,96 – 1,14)	14.04.15	5	1,07	1,04
		27.07.15	5	1,07	1,08
Equalis INR calibrator High	<b>3,14</b> (2,57 – 3,71)	14.04.15	5	3,09	3,12
		27.07.15	5	3,20	3,22
Equalis INR control	<b>2,48</b> (2,09 – 2,87)	14.04.15	5	2,40	2,43
		27.07.15	5	2,49	2,49

**Table 6.** DEKS PT (INR) calibrators measured on the comparison method

Material	Assigned value, PT (INR) (uncertainty)	Date	n	Mean value, PT (INR) instrument 1	Mean value, PT (INR) instrument 2
DEKS INR calibrator Normal	<b>1,00</b> (0,98 – 1,03)	14.04.15	5	1,02	0,99
		13.05.15	5	0,99	1,01
		27.07.15	5	1,02	1,04
DEKS INR calibrator Therapeutic	<b>2,26</b> (2,19 – 2,33)	14.04.15	5	2,23	2,22
		13.05.15	5	2,24	2,24
		27.07.15	5	2,25	2,27
DEKS INR calibrator High	<b>3,74</b> (3,59 – 3,89)	14.04.15	5	3,81	3,86
		13.05.15	5	3,83	3,82
		27.07.15	5	3,89	3,91

Results achieved for external quality control material from Noklus/NKK EQA-scheme in January, May and August 2015 show that the deviations from the assigned value for the three surveys were (0,00); (- 0,02); (-0,02) INR at a normal level and (-0,08); (-0,03); (0,00) INR at a therapeutic level respectively. Results of the Noklus/NKK control material, which also were analysed at different occasions during the evaluation, were within the acceptable limits for the control material (data not shown).

### *Discussion*

The results from the comparison method were in agreement with both the Equalis calibrators and the DEKS calibrators, and there was good agreement between the two STAR-instruments representing the comparison method. The results from the EQA-scheme showed that the comparison method was in agreement with the other hospital laboratories (n=67–69) using PT (INR) calibrators from Equalis.

## **6.3. Analytical quality of the microINR system under optimal conditions**

The results below reflect the analytical quality of the microINR system under optimal conditions. The results documents the quality of the system under conditions as favourable as possible for achieving good analytical quality.

### **6.3.1. Internal analytical quality control**

All the the results from the internal analytical quality control (microINR Easy Control), one level, were within the allowable control limits. The reproducibility achieved with the internal analytical quality control samples was 8,1% (n=36). Raw data is attached for the requesting company only, attachment 7.

### **6.3.2. The precision of the microINR system**

The measurements with the microINR system were performed with two capillary samples from each patient. The duplicate results were checked to meet the imposed condition for using formula 1 in attachment 5 (data not shown). For PT (INR) values > 2,5 no systematic deviation between the paired measurements was found. Unexpectedly, there was a small systematic difference pointed out in the low PT (INR) level. The second result of the duplicate measurements in this PT (INR) level was in average 0,1 INR lower than the first measurement. A smaller, but still statistic significant systematic difference, was also found for the low PT (INR) results in one of the

PHCCs (see 6.4.2). No explanation for these systematic differences has been found. The sampling procedure and sample handling, as well as the measurement procedure, were identical for all samples and measurements throughout the evaluation. In a simulated set of data, the average systematic difference was removed, and the CV was recalculated. As a consequence, no result for the repeatability at PT (INR) level  $<2,5$  is given in table 7, but an approximate range for the repeatability at PT (INR) level  $<2,5$  is discussed.

The precision is presented as the repeatability. The CV with a 90% CI is shown in table 7. The results are sorted and divided into two concentration levels according to the mean of the results of the microINR system. Raw data is attached for the requesting company only, attachment 8.

**Table 7.** Repeatability, PT (INR) capillary samples microINR. Results achieved under optimal conditions

PT (INR) level microINR	n	Excluded results	Mean value (interval) PT (INR)	CV (90% CI) %
$<2,5$	34	0	2,0 (1,00 – 2,45)	*
$\geq 2,5$	64	0	3,0 (2,50 – 4,70)	6,0 (5,3 – 7,0)

An account of the number of samples is given in section 6.1.

\* If the repeatability is calculated without taking into consideration the systematic difference pointed out between the paired measurements in this level, the CV is 5,6. If the average systematic difference is removed from all paired results in this level, the estimated CV becomes 4,1.

### Discussion

At PT (INR) level  $<2,5$  the CV was estimated to be somewhere between 4,1 and 5,6%. Because of increased uncertainty in this estimate, the confidence interval for the estimated values is not given. The CV was anyway not significantly below 5,0%. At PT (INR) level  $\geq 2,5$  the CV was 6,0%, with the confidence interval above 5%.

### Conclusion

The repeatability under optimal conditions at PT (INR) level  $<2,5$  was somewhere between 4,1 and 5,6%, and most likely fulfilled the quality goal of  $\leq 5\%$ . At PT (INR)  $\geq 2,5$  the quality goal for the repeatability was not fulfilled.

#### 6.3.3. The trueness of the microINR system

The mean deviation (bias) of the microINR system results from the comparison method was calculated. The bias is presented with a 95% CI in table 8. The results are sorted and divided into two concentration levels according to the mean results of the comparison method. Raw data is attached for the requesting company only, attachment 6 and 8.

**Table 8.** Bias, PT (INR) capillary samples microINR. Results achieved under optimal conditions

PT (INR) level Comparison method	n	Excluded results	Mean value Comparison method, PT (INR)	Mean value PT (INR)	Bias (95% CI), INR	Bias , %
<2,5	38	1*	2,1	2,1	+0,06 (0,002 – +0,13)	3,1
≥2,5	57	2*	3,0	3,0	-0,04 (-0,11 – +0,02)	-1,4


\* The given numbers of results (n) are counted before the exclusion of results. Mean and bias are calculated after the exclusion of results. ID 9, ID 24 and ID 64 are statistical outliers according to Burnett's model [15] and therefore excluded. An account of the number of samples is given in section 6.1.

### Discussion

For PT (INR) level <2,5 a small, but statistically significant bias was shown. The microINR system gave results 3,1% higher than the comparison method with an average mean bias of +0,06 INR. For PT (INR) level ≥2,5 there was no statistically significant bias pointed out, and the microINR results were in agreement with the comparison method.

#### 6.3.4. The accuracy of the microINR system

To evaluate the accuracy of PT (INR) results from the microINR system, the agreement between microINR and the comparison method is illustrated in a difference plot (figure 2). The limits for the allowable deviation according to the quality goal ( $\pm 20\%$ ) are shown with stippled lines. The plot illustrates both random and systematic errors, reflecting the total measuring error in the microINR results. The total number of results from microINR are included in the plot. Raw data is attached for the requesting company only, attachment 6 and 8.



**Figure 2.** Accuracy of PT (INR) on the microINR system under optimal conditions. The x-axis represents the mean PT (INR) result of the corresponding sample of the comparison method. The y-axis represents the deviation in PT (INR) of the first capillary sample measurement on microINR from the mean result of the comparison method. Different lots are illustrated as Lot 47170 (●), lot 50020 (◆) and lot 50120 (\*). Stippled lines represent allowable deviation limits of  $\pm 20\%$ . Number of results (n) = 95. An account of the number of samples, and excluded and missing results, is given in section 6.1.

### *Discussion*

In (PT) INR level  $< 2,5$  the difference plot shows more results above zero than below, which correspond to the calculated bias in 6.3.3. In the higher PT (INR) level, there is a tendency for the microINR system giving results below the comparison method.

In figure 2, three of 95 results achieved under optimal conditions were outside the allowable deviation limits of  $\pm 20\%$ . Two of these results (2%) deviate  $> \pm 25\%$ . There is no comments registered regarding these results. The share of results within the limits was 97%.

### *Conclusion*

The quality goal for accuracy under optimal conditions was fulfilled.

#### **6.3.5. Bias with three lots of test chips**

As can be seen in figure 2, there were no differences in the distribution of the results of the three lots. Separate lot calculations were not performed.

#### **6.4. Analytical quality of the microINR system achieved by intended users**

The results below reflect the analytical quality of the microINR system under real conditions in the hands of intended users. The results may deviate from the results achieved under optimal conditions.

##### **6.4.1. Internal analytical quality control**

A total of 23 out of 24 results from the internal analytical quality control (microINR Easy Control), one level, were within the allowable control limits. The reproducibility achieved with the internal analytical quality control samples was 8,6% (n=12) in PHCC1 and 7,9% (n=12) in PHCC2. Raw data is attached for the requesting company only, attachment 9.

##### **6.4.2. The precision of the microINR system**

The measurements with the microINR system were performed with two capillary samples from each patient. The duplicate results were checked to meet the imposed condition for using formula 1 in attachment 5 (data not shown). For PT (INR) values  $> 2,5$ , no systematic deviation between the paired measurements was found. There was a small systematic difference pointed out in the low PT (INR) level in the results from PHCC2. The second result of the duplicate measurements in this PT (INR) level, was in average 0,06 PT (INR) lower than the first measurement. This phenomenon was also found for the low PT (INR) results achieved under optimal conditions (see 6.3.2). No explanation for these systematic differences has been found. The sampling procedure and sample handling, as well as the measurement procedure, were identical for all samples and measurements throughout the evaluation. The systematic differences pointed out lead to an overestimation of the CV. In a simulated set of data, the average systematic difference was removed, and the CV was recalculated. As a consequence, no result for the repeatability at PT (INR) level  $< 2,5$  achieved by PHCC2 is given in table 9, but an approximate range for the repeatability at PT (INR) level  $< 2,5$  is discussed.

The precision is presented as the repeatability. The CV with a 90% CI is shown in table 9. The results are sorted and divided into two concentration levels according to the mean of the results of microINR. Raw data is attached for the requesting company only, attachment 10.

**Table 9.** Repeatability, PT (INR) capillary samples microINR. Results achieved by intended users

PT (INR) level microINR	n	Excluded results	Mean value (interval) PT (INR)	CV (90% CI) %
<i>PHCC1</i>				
< 2,5	13	0	2,2 (1,85 – 2,45)	5,0 (3,7 – 7,8)
≥2,5	18	0	2,9 (2,50 – 3,65)	6,1 (4,8 – 8,6)
<i>PHCC2</i>				
< 2,5	24	0	2,1 (1,35 – 2,45)	*
≥2,5	24	0	3,1 (2,50 – 7,65)	6,3 (5,1 – 8,3)

An account of the number of samples is given in section 6.1.

\* If the repeatability is calculated without taking into consideration the systematic difference pointed out between the paired measurements in this level from PHCC2, the CV is 4,7. If the average systematic difference is removed from all paired results in this level, the estimated CV becomes 4,2.

### *Discussion*

The CV achieved by PHCC1 at PT (INR) level <2,5 was 5,0%, and the upper CI value was above the quality goal. For PHCC2 the CV at this PT (INR) level, was estimated to be somewhere between 4,2 and 4,7%. Because of increased uncertainty in this estimate, the confidence interval for the estimated values is not given, but the estimated CV was not significantly below 5%. At PT (INR) level ≥2,5 the CV was 6,1% and 6,3% at the two PHCCs, respectively. This is higher than the quality goal, but for PHCC1 not statistically significant higher.

### *Conclusion*

The repeatability achieved by intended users at PT (INR) level <2,5 most likely fulfilled the quality goal of ≤5%. The repeatability achieved by intended users at PT (INR) level ≥2,5 did most likely not fulfill the quality goal for PHCC1, and did not fulfill it for PHCC2.

### 6.4.3. The bias of the microINR system

The mean deviation (bias) of the microINR system results from the comparison method was calculated. The bias is presented with a 95% CI in table 10. The results are sorted and divided into two concentration levels according to the mean results of the comparison method. Raw data is attached for the requesting company only, attachment 6 and 10.

**Table 10.** Bias, PT (INR) capillary samples microINR. Results achieved by intended users

PT (INR) level Comparison method	n	Excluded results	Mean value Comparison method, PT (INR)	Mean value PT (INR)	Bias (95% CI), INR	Bias, %
<i>PHCC1</i>						
< 2,5	19	0	2,18	2,26	+0,08 (-0,03 – +0,20)	3,8
≥2,5	17	0	3,06	2,94	-0,11 (-0,25 – +0,02)	-3,8
<i>PHCC2</i>						
< 2,5	22	0	2,12	2,15	+0,03 (-0,05 – +0,11)	1,5
≥2,5	19	0	3,24	3,16	-0,08 (-0,25 – +0,09)	-2,6

An account of the number of samples is given in section 6.1.


### Discussion

For both PT (INR) levels no statistically significant bias was pointed out on the results achieved by the intended users. The microINR results were in agreement with the results from the comparison method. The results were in accordance with the results achieved under optimal conditions.

### 6.4.4. The accuracy of the microINR system

To evaluate the accuracy of PT (INR) results from the microINR system, the agreement between microINR and the comparison method is illustrated in a difference plot (figure 3). The limits for the allowable deviation according to the quality goal ( $\pm 20\%$ ), are shown with stippled lines. The plot illustrates both random and systematic errors, reflecting the total measuring error in the microINR results. The total number of results from microINR are included in the plot. Raw data is attached for the requesting company only, attachment 6 and 10.





**Figure 3.** Accuracy of PT (INR) on the microINR system achieved by intended users. The x-axis represents the mean PT (INR) result of the corresponding sample of the comparison method. The y-axis represents the deviation in PT (INR) of the first capillary samples measurement on microINR from the mean result of the comparison method. The results from PHCC1 are represented with the symbol (●) and results from PHCC2 with the symbol (◆). Stippled lines represent allowable deviation limits of  $\pm 20\%$ . Number of results ( $n$ ) = 77. One result at 8.5 INR, is not shown in the figure, but illustrated as  $\rightarrow$ . The result was within the allowable deviation limits. An account of the number of samples, and excluded and missing results, is given in section 6.1.

### Discussion

The difference plot shows a negative bias in the higher PT (INR) level, corresponding to the calculated bias in 6.4.3. In the higher PT (INR) level, there is a tendency for the microINR system giving results below the comparison method. The results from the PHCCs are in agreement with the results achieved under optimal conditions.

In figure 3, four of 77 results achieved by intended users were outside the allowable deviation limits of  $\pm 20\%$ . Two of these results (2,6%) deviate  $> \pm 25\%$ . No comments attached to the results can explain the deviations. The share of results within the limits was 95%.

### Conclusion

The quality goal for accuracy achieved by intended users was fulfilled.

## 6.5. Evaluation of user-friendliness

### 6.5.1. *Questionnaire to the evaluators*

The most important response regarding user-friendliness comes from the intended users themselves. The end-users often emphasize other aspects than those pointed out by more extensively trained laboratory personnel.

At the end of the evaluation period, the intended user fills in a questionnaire about the user-friendliness of the instrument. SKUP has prepared detailed instructions for this.

The questionnaire is divided into four subareas:

Table A) Rating of the information in the manual / insert / quick guide

Table B) Rating of operation facilities. Is the system easy to handle?

Table C) Rating of time factors for the preparation and the measurement

Table D) Rating of performing internal and external analytical quality control

The intended users fill in table A and B. SKUP fills in table C and D and in addition, topics marked with grey colour in table A and B.

In the tables, the first column shows what is up for consideration. The second column in table A and B shows the rating by the individual users at the evaluation sites. The last three columns show the rating options. The overall ratings from all the evaluating sites are marked in coloured and bold text. The last row in each table summarises the total rating in the table. The total rating is an overall assessment by SKUP of the described property, and not necessarily the arithmetic mean of the rating in the rows. Consequently, a single poor rating can justify an overall poor rating, if this property seriously influences on the user-friendliness of the system.

Unsatisfactory and intermediate ratings are marked with a number and explained below the tables. The intermediate category covers neutral ratings assessed as neither good nor bad.

An assessment of the user-friendliness is subjective, and the topics in the questionnaire may be emphasised differently by different users. The assessment can therefore vary between different persons and between the countries. This will be discussed and taken into account in the overall assessment of the user-friendliness.

#### *Comment*

In this evaluation, the user-friendliness was assessed by PHCC1 (the opinion of one BLS and three health secretaries) and PHCC2 (the opinion of two health secretaries and one medical secretary). PHCC2 did not evaluate table A.

**Table A.** Rating of the information in the manual

Topic	Rating	Assessment	Assessment	Assessment
General impression	I <sup>1</sup>	Satisfactory	Intermediate	Unsatisfactory
Table of contents	S	Satisfactory	Intermediate	Unsatisfactory
Preparations / Pre-analytic procedure	S	Satisfactory	Intermediate	Unsatisfactory
Specimen collection	I <sup>2</sup>	Satisfactory	Intermediate	Unsatisfactory
Measurement procedure	I <sup>2</sup>	Satisfactory	Intermediate	Unsatisfactory
Reading of result	S	Satisfactory	Intermediate	Unsatisfactory
Description of the sources of error	S	Satisfactory	Intermediate	Unsatisfactory
Help for troubleshooting	I <sup>2</sup>	Satisfactory	Intermediate	Unsatisfactory
Readability / Clarity of presentation	U <sup>3</sup>	Satisfactory	Intermediate	Unsatisfactory
Keyword index*		Satisfactory	Intermediate	Unsatisfactory
Measurement principle	I <sup>4</sup>	Satisfactory	Intermediate	Unsatisfactory
Available insert in Danish, Norwegian, Swedish	S	Satisfactory	Intermediate	Unsatisfactory
<b>Total rating by SKUP</b>			<b>Intermediate</b>	

\* Not rated in this evaluation due to the small size of the manual

<sup>1</sup>In general there are too many difficult words in the manual

<sup>2</sup>Some of the instructions are ambiguous

<sup>3</sup>Small print

<sup>4</sup>Explanation of the measurement principal is affected by poor translation into Norwegian

**Table B.** Rating of operation facilities

Topic	Rating	Assessment	Assessment	Assessment
To prepare the test / instrument	<b>I<sup>1</sup>, U<sup>1</sup></b>	Satisfactory	<b>Intermediate</b>	<b>Unsatisfactory</b>
To prepare the sample	<b>S, U<sup>2</sup></b>	<b>Satisfactory</b>	Intermediate	<b>Unsatisfactory</b>
Application of specimen	<b>U<sup>3</sup>, U<sup>3</sup></b>	Satisfactory	Intermediate	<b>Unsatisfactory</b>
Specimen volume	<b>S, S</b>	<b>Satisfactory</b>	Intermediate	Unsatisfactory
Number of procedure step	<b>S, I<sup>4</sup></b>	<b>Satisfactory</b>	<b>Intermediate</b>	Unsatisfactory
Instrument / test design	<b>I<sup>5</sup>, U<sup>6</sup></b>	Satisfactory	<b>Intermediate</b>	<b>Unsatisfactory</b>
Reading of the test result	<b>S, S</b>	<b>Easy</b>	Intermediate	Difficult
Sources of errors	<b>U<sup>7</sup>, -</b>	Satisfactory	Intermediate	<b>Unsatisfactory</b>
Cleaning / Maintenance	<b>S, -</b>	<b>Satisfactory</b>	Intermediate	Unsatisfactory
Hygiene, when using the test	<b>S, U<sup>8</sup></b>	<b>Satisfactory</b>	Intermediate	<b>Unsatisfactory</b>
Size and weight of package	<b>S, S</b>	<b>Satisfactory</b>	Intermediate	Unsatisfactory
Storage conditions for tests, unopened package*	<b>S</b>	<b>+15 to +30°C</b>	+2 to +8°C	-20°C
Storage conditions for tests, opened package**		+15 to +30°C	+2 to +8°C	-20°C
Environmental aspects: waste handling	<b>S</b>	<b>No precautions</b>	Sorted waste	Special precautions
Intended users	<b>S</b>	<b>Health care personnel or patients</b>	Laboratory experience	Biomedical laboratory scientists

**Total rating by SKUP****Unsatisfactory**

\*Storage temperature for the microINR chips is +2 to +25°C

\*\*Not rated in this evaluation due to the single-pack concept of the microINR chip

<sup>1</sup>The period of time from the meter is switched on to the blood can be placed on the test chip, is too long (comment from SKUP: iLine Microsystems S.L. informs that the start-up procedure takes from 50-70 seconds under standard conditions.)

<sup>2</sup>The meter could not be moved (comment from SKUP: as instructed by iLine Microsystems S.L. and Orion Diagnostica during the training, see 5.2 and 5.4.1) and the finger could not touch the test chip when applying the sample. The system is too sensitive and has too many sources of error.

<sup>3</sup>The test chip is too sensitive to the way the sample is applied

<sup>4</sup>Too many details to take care of when the sample is to be placed on the microINR chip

<sup>5</sup>Good-sized meter and clear display, but difficult to analyse the internal quality control material

<sup>6</sup>When the test chip is inserted in the meter and the sample is about to be placed on the chip, it becomes too little space to do it properly. It would be better if the meter could be lifted towards the finger

<sup>7</sup>Many error codes due to the sensitivity to the way the sample is applied

<sup>8</sup>It is possible to spill blood when a used test chip is removed from the meter

Additional negative comment: The battery needs to be recharged more often than specified in the manual.

**Table C.** Rating of time factors (filled in by SKUP)

Topic	Assessment	Assessment	Assessment
Required training time	<2 hours	2 to 8 hours	>8 hours
Durations of preparations / Pre-analytical time	<6 min.	6 to 10 min.	>10 min.
Duration of analysis	<10 min.	10 to 20 min.	>20 min.
Stability of test, unopened package	>5 months	3 to 5 months	<3 months
Stability of test, opened package*	>30 days	14 to 30 days	<14 days
Stability of quality control material, unopened**	>5 months	3 to 5 months	<3 months
Stability of quality control material, opened***	>6 days or disposable	2 to 6 days	≤1 day

**Total rating by SKUP**

**Satisfactory**

\*Not rated in this evaluation due to the single-pack concept of the microINR chip

\*\*The stability of the internal analytical control material microINR Easy Control is >5 months if the control material is stored at +2 to +8°C

\*\*\*Not rated in this evaluation due to the fact that clotting will start immediately after reconstitution

Additional negative comments: There is no information about the storage conditions of the chip in the manual, only in the package insert of the chip.

**Table D.** Rating of analytical quality control (filled in by SKUP)

Topic	Assessment	Assessment	Assessment
Reading of the internal quality control	Satisfactory	Intermediate	Unsatisfactory
Usefulness of the internal quality control	Satisfactory	Intermediate <sup>1</sup>	Unsatisfactory
External quality control	Satisfactory	Intermediate	Unsatisfactory
<b>Total rating by SKUP</b>	<b>Satisfactory</b>		

<sup>1</sup>The imprecision achieved with the internal analytical control material microINR Easy Control is above the imprecision of the patient samples, i.e. less possibilities to discover errors in the analytical system

### 6.5.2. Assessment of the user-friendliness

#### *Assessment of the information in the manual (table A)*

The manual was assessed as intermediate and most of the comments are related to difficult words and poor translation.

#### *Assessment of the operation facilities (table B)*

The operation facilities were assessed to unsatisfactory based on the consistent feedback of the sources of error associated with the application of the sample. In total 4,3% of the measurements failed due to errors.

#### *Assessment of time factors (table C)*

The time factors were assessed as satisfactory.

#### *Assessment of analytical quality control possibilities (table D)*

The analytical quality control possibilities were assessed as satisfactory.

#### *Conclusion*

In total, the user-friendliness of the microINR system and its manual was rated as intermediate. This does not fulfill the quality goal for user-friendliness.

## 7. References

1. Christensen N.G., Monsen G. & Sandberg S. *Utprøving av analyseinstrumenter*, 1997. Alma Mater Forlag.
2. ISO/FDIS 17593:2007, Clinical laboratory testing and in vitro medical devices – Requirements for in vitro monitoring systems for self-testing of oral anticoagulant therapy.
3. Hyltoft Petersen P. *et al.* Combination of analytical quality specifications based on biological within- and between-subject variation. *Ann Clin Biochem* 2002; **39** (6): 543 – 550.
4. Fraser C.G. & Hyltoft Petersen P. Quality goals in external quality assessment are best based on biology. *Scand J Clin Lab Invest* 1993; **53** suppl 212. Chapter I. Quality planning.
5. Ricos C. *et al.* Current databases on biological variation: pros, cons and progress. *Scand J Clin Lab Invest* 1999; **66** (4): 337 – 349.
6. Kjeldsen J., Lassen J.F., Hyltoft Petersen P. & Brandslund I. Biological variation of international normalized ratio for prothrombin times, and consequences in monitoring oral anticoagulant therapy: computer simulation of serial measurements with goalsetting for analytical quality. *Clin Chem* 1997; **43** (11): 2175 – 2182.
7. Kvalitetskrav og kvalitetsvurdering for hyppigt udførte klinisk biokemiske og klinisk mikrobiologiske analyser i almen praksis. Konsensus dokument udarbejdet af Laboratorieudvalget under Sygesikringens og PLO's Faglige Udvalg vedr. Almen Praksis i samarbejde med DEKS og Dansk Selskab for Klinisk Biokemi's Videnskabelige udvalg. Nov 2003. [www.skup.nu](http://www.skup.nu) (menu The SKUP evaluations – Quality goals).
8. Kvalitetssikring og kvalitetskrav til laboratoriemedicinske aktiviteter i almen praksis. Udarbejdet af Regionernes Lønnings- og Takstnævn (RTLN) og Praktiserende Lægers Organisation (PLO). 2010. [www.skup.nu](http://www.skup.nu) (menu The SKUP evaluation – Quality goals).
9. <http://www.ifcc.org/ifcc-scientific-division/sd-committees/c-npu/npusearch/>
10. Van den Besselaar A.M. Multicentre study of replacement of the international reference preparation for thromboplastin rabbit plain. *Thromb Haemost* 1993; **70**: 794 – 799.
11. Van den Besselaar A.M. & Houdijk W.P. Use of lyophilized calibrant plasmas for simplified international normalized ratio determination with a human tissue factor thromboplastin reagent derived from cultured human cells. *Clin Chem* 2003; **49** (12): 2006 – 2011.
12. WHO. Expert committee on biological standardization – Requirements for thromboplastins and plasma used to control oral anticoagulant therapy. World Health Organization, Geneva Technical Report Series 889, 48th Report 1999.
13. Hillarp A. *et al.* Local INR calibration of the Owren type prothrombin assay greatly improves the intra- and interlaboratory variation. *Thromb Haemost* 2004; **91**: 3300 – 3307.
14. Arbetsbeskrivning A147, ver 1.0, 2013. Arbetsbeskrivning för Equalis och Expertgrupp vid åsättande av INR-värden till kalibratorer och kontrollmaterial för bestämning av protrombinkomplex enligt Owren. Equalis, Uppsala.

15. Burnett RW. Accurate estimation of standard deviations for quantitative methods used in clinical chemistry. *Clinical Chemistry* 1975; **21** (13): 1935 – 1938.



## **Attachments**

1. The organisation of SKUP
2. Facts about the microINR system
3. Information about manufacturer, retailers and marketing
4. Product specifications for this evaluation, microINR system
5. Statistical expressions and calculations
6. Raw data PT (INR), results from the comparison method
7. Raw data PT (INR), internal analytical quality control microINR, optimal conditions
8. Raw data PT (INR), microINR results, optimal conditions
9. Raw data PT (INR), internal analytical quality control, microINR, intended users
10. Raw data PT (INR) microINR results, intended users
11. “SKUP-info”. Summary for primary health care (in Norwegian))
12. List of previous SKUP evaluations
13. Comments from iLine Microsystems S.L.

Attachments with raw data are included only in the copy to iLine Microsystems S.L.

## The organisation of SKUP

*Scandinavian evaluation of laboratory equipment for primary health care, SKUP*, is a co-operative commitment of Noklus<sup>1</sup> in Norway, Denmark<sup>2</sup>, and Equalis<sup>3</sup> in Sweden. SKUP was established in 1997 at the initiative of laboratory medicine professionals in the three countries. SKUP is led by a Scandinavian *steering committee* and the secretariat is located at Noklus in Bergen, Norway.

*The purpose of SKUP* is to improve the quality of near patient testing in Scandinavia by providing objective and supplier-independent information on analytical quality and user-friendliness of laboratory equipment. This information is generated by organising SKUP *evaluations*.

SKUP offers manufacturers and suppliers evaluations of equipment for primary health care and also of devices for self-monitoring. Provided the equipment is not launched onto the Scandinavian market, it is possible to have a confidential pre-marketing evaluation. The company requesting the evaluation pays the actual testing costs and receives in return an impartial evaluation.

There are *general guidelines* for all SKUP evaluations and for each evaluation, a specific *SKUP protocol* is worked out in co-operation with the manufacturer or their representatives. SKUP signs *contracts* with the requesting company and the evaluating laboratories. A *complete evaluation* requires one part performed by experienced laboratory personnel as well as one part performed by the intended users.

Each evaluation is presented in a *SKUP report* to which a unique *report code* is assigned. The code is composed of the acronym SKUP, the year and a serial number. A report code, followed by an asterisk (\*), indicates a special evaluation, not complete according to the guidelines, e.g. the part performed by the intended users was not included in the protocol. If suppliers use the SKUP name in marketing, they have to refer to [www.skup.nu](http://www.skup.nu) and to the report code in question. For this purpose the company can use a logotype available from SKUP containing the report code.

SKUP reports are published at [www.skup.nu](http://www.skup.nu).

---

<sup>1</sup> Noklus (Norwegian Quality Improvement of Primary Care Laboratories) is an organisation founded by Kvalitetsforbedringsfond III (Quality Improvement Fund III), which is established by The Norwegian Medical Association and the Norwegian Government. Noklus is professionally linked to “Seksjon for Allmenntmedisin” (Section for General Practice) at the University of Bergen, Norway.

<sup>2</sup> SKUP in Denmark is placed in Nordsjællands Hospital. Currently SKUP in Denmark is out of operation due to lack of funding.

<sup>3</sup> Equalis AB (External quality assurance in laboratory medicine in Sweden) is a limited company in Uppsala, Sweden, owned by “Sveriges Kommuner och Landsting” (Swedish Association of Local Authorities and Regions), “Svenska Läkaresällskapet” (Swedish Society of Medicine) and IBL (Swedish Institute of Biomedical Laboratory Science).

## Facts about the microINR measurement system

Parts of this form are filled in by iLine Microsystems S.L.

**Table 1. Basic facts**

Name of the measurement system:	microINR system (microINR Meter and microINR Chips)
Dimensions and weight:	Width: 65 mm Depth: 35 mm Height: 119 mm Weight: 205 gr
Components of the measurement system:	microINR Meter and microINR Chips
Measurand:	PT (INR)
Sample material:	Fresh capillary blood
Sample volume:	3µl
Measuring principle:	Microfluidics & Machine Vision System
Traceability:	All lots of microINR Chips are traceable to the International Reference Preparation, (rTF/09+tilt tube).
Calibration:	The lot calibration is printed on the Datamatrix of the Chip. This is read by the Meter and applied in every test.
Measuring range:	0.8 - 8 INR
Linearity:	
Measurement duration:	
Operating conditions:	Temperature from 15°C to 35°C. Relative humidity below 80%
Electrical power supply:	Rechargeable battery
Recommended regular maintenance:	Clean / Disinfect the microINR Meter when it is visibly dirty.
Package contents:	Contains 1 microINR meter, 1 carrier case, 1 charger, 1 USB wire, 1 lancing device, 10 lancets and 1 CD with data extraction software and instructions. microINR Chips sold separately.
Necessary equipment not included in the package:	None

**Table 2. Post analytical traceability**

Is input of patient identification possible?	Yes, manually
Is input of operator identification possible?	No
Can the instrument be connected to a bar-code reader?	No
Can the instrument be connected to a printer?	No
What can be printed?	N/A

Can the instrument be connected to a PC?	Yes
Can the instrument communicate with LIS (Laboratory Information System)? If yes, is the communication bidirectional?	No. But yes thought a middle ware that has been programmed to recognize microINR transferred data. The communication will be unidirectional.
What is the storage capacity of the instrument and what is stored in the instrument?	199 results including errors
Is it possible to trace/search for measurement results?	No

**Table 3. Facts about the reagent/test strips/test cassettes**

Name of the reagent/test strips/test cassettes:	microINR Chips
Stability in unopened sealed vial:	1 year at room temperature (both box and blister)
Stability in opened vial:	6 hours
Package contents:	Contains 25 microINR Chips to be used with the microINR meter and 1 set of instructions.

**Table 4. Quality control**

Electronic self check:	Yes
Recommended control materials and volume:	microINR EasyControl
Stability in unopened sealed vial:	6 months (both box and sealed vial)
Stability in opened vial:	14 min once reconstituted.
Package contents:	Contains 5 plasma vials, 5 calcium solution vials, 5 capillary pipettes and 1 set of instructions

## Information about manufacturer, retailers and marketing

**Table 1. Marketing information**

Manufacturer:	iLine Microsystems S.L.
Retailers in Scandinavia:	<u>Denmark: Orion Diagnostica oy</u> <u>Norway: Orion Diagnostica as</u> <u>Sweden: Orion Diagnostica AB</u>
In which countries is the system marketed:	Globally
Date for start of marketing the system in Scandinavia:	December 2012
Date for CE-marking:	February 2011
In which Scandinavian languages is the manual available:	All

## Product specifications for this evaluation, the microINR system

### *microINR serial numbers*

<b>Instrument</b>	<b>Serial number</b>	<b>Used by</b>
microINR	1473260128001	St. Olavs Hospital
microINR	1556240436001	PHCC1
microINR	1587570016001	PHCC2
microINR	1473260187001	extra

### *microINR chips*

<b>PT test chips</b>	<b>Lot number</b>	<b>Expiry date</b>	<b>Used by</b>
Test chip lot A	47170	2015-10	St. Olavs Hospital, PHCC1, PHCC2
Test chip lot B	50020	2016-01	St. Olavs Hospital, PHCC1, PHCC2
Test chip lot C	50120	2016-01	St. Olavs Hospital, PHCC1, PHCC2

### *Other equipment used in the evaluation*

<b>Other equipment</b>	<b>Lot number</b>	<b>Expiry date</b>	<b>Used by</b>
Vacurette 3,2% sodium citrate tube			St.Olavs, PHCC1, PHCC2
microINR EasyControl	150317	2015-09-17	St.Olavs, PHCC1, PHCC2
Cutisoft Wipes skin clean			St.Olavs, PHCC1, PHCC2
Accu-Chek Safe-T-Pro Plus lancet	Y080020	2018-09	St.Olavs, PHCC1, PHCC2

## Statistical expressions and calculations

This chapter with standardised text deals with the statistical expressions and calculations used by SKUP. The statistical calculations will change according to the type of evaluation. The descriptions in this document are valid for evaluations of quantitative methods with results on the ratio scale.

## Statistical terms and expressions

The definitions in this section come from the ISO/IEC Guide 99; International Vocabulary of Metrology, VIM [a].

### Precision

*Definition: Precision is the closeness of agreement between measured quantity values obtained by replicate measurements on the same or similar objects under stated specified conditions.*

Precision is measured as *imprecision*. Precision is descriptive in general terms (good, poor e.g.), whereas the imprecision is expressed by means of the standard deviation (SD) or coefficient of variation (CV). SD is reported in the same PT (INR) as the analytical result. CV is usually reported in percent.

To be able to interpret an assessment of precision, the precision conditions must be defined. *Repeatability* is the precision of consecutive measurements of the same component carried out under identical measuring conditions (within the measuring series).

*Reproducibility* is the precision of discontinuous measurements of the same component carried out under changing measuring conditions over time.

### Trueness

*Definition: Trueness is the closeness of agreement between the average of an infinite number of replicate measured quantity values and a reference quantity value.*

Trueness is inversely related to systematic measurement error. Trueness is measured as *bias*. Trueness is descriptive in general terms (good, poor e.g.), whereas the bias is reported in the same PT (INR) as the analytical result or in percent.

### Accuracy

*Definition: Accuracy is the closeness of agreement between a measured quantity value and the true quantity value of a measurand.*

Accuracy is not a quantity and cannot be expressed numerically. A measurement is said to be more accurate when it offers a smaller measurement error. Accuracy can be illustrated in a difference-plot. Accuracy is descriptive in general terms (good, poor e.g.).

- a. ISO/IEC Guide 99:2007, International vocabulary of metrology – Basic and general concepts and associated terms, VIM, 3<sup>rd</sup> edition, JCGM 200:2008.

## Statistical calculations

### Statistical outliers

The criterion promoted by Burnett [b] is used for the detection of outliers. The model takes into consideration the number of observations together with the statistical significance level for the test. The significance level is set to 5%. The segregation of outliers is made with repeated truncations, and all results are checked. Where the results are classified according to different concentration levels, the outlier-testing is carried out at each level separately. Statistical outliers are excluded from the calculations.

### Calculation of imprecision

The precision of the evaluated method is assessed by use of paired measurements of genuine patient sample material. The results are usually divided into three concentration levels, and the estimate of imprecision is calculated for each level separately, using the following formula [c,d]:

$$SD = \sqrt{\frac{\sum d^2}{2n}} \quad \begin{array}{l} d = \text{difference between two paired measurements} \\ n = \text{number of differences} \end{array}$$

This formula is used when the standard deviation can be assumed reasonable constant across the concentration interval. If the coefficient of variation is more constant across the concentration interval, the following formula is preferred:

$$CV = \sqrt{\frac{\sum (d/m)^2}{2n}} \quad \begin{array}{l} m = \text{mean of paired measurements} \end{array} \quad (\text{formula 2})$$

The two formulas are based on the differences between paired measurements. The calculated standard deviation or CV is still a measure of the imprecision of single values. The imposed condition for using the formulas is that there is no systematic difference between the 1<sup>st</sup> and the 2<sup>nd</sup> measurement of the pairs. The CV is given with a 90% confidence interval.

### Calculation of bias

The mean deviation (bias) at different concentration levels is calculated based on results achieved under optimal measuring conditions. A paired t-test is used with the mean values of the duplicate results on the comparison method and the mean values of the duplicate results on the evaluated method. The mean difference is shown with a 95% confidence interval.

### Assessment of accuracy

The agreement between the evaluated method and the comparison method is illustrated in a difference-plot. The x-axis represents the mean value of the duplicate results on the comparison method. The y-axis shows the difference between the first measurement on the evaluated method and the mean value of the duplicate results on the comparison method. The number of results within the quality goal limits is counted and assessed.

- b. Burnett RW. Accurate estimation of standard deviations for quantitative methods used in clinical chemistry. *Clinical Chemistry* 1975; **21** (13): 1935 – 1938.
- c. Saunders E. *Tietz textbook of clinical chemistry and molecular diagnostics*, 2006. Chapter 14, Linnet K., Boyd J. Selection and analytical evaluation of methods – with statistical techniques. Elsevier Saunders ISBN 0-7216-0189-8.
- d. Fraser C.G. *Biological variation: From principles to practice*, 2006. Chapter 1, The Nature of Biological Variation. AACC Press ISBN 1-890883-49-2.



# SKUP-info



## Sammendrag fra en utprøving i regi av SKUP

microINR for måling av PT-INR

**Produsent:** iLine Microsystems S.L.

**Norsk forhandler:** Orion Diagnostica as

### Konklusjon

For PT-INR resultat under 2,5 ble kvalitetsmålet for presisjon mest sannsynlig oppfylt. For PT-INR resultat over 2,5 ble kvalitetsmålet for presisjon ikke oppfylt. Kvalitetsmålet for nøyaktighet ble oppfylt. Kvalitetsmålet for brukervennlighet ble ikke oppfylt.

### Bakgrunn

microINR er et bærbart koagulometer for måling av protrombintid, PT-INR. Systemet er beregnet for profesjonell bruk og til egenmåling og egenkontroll av pasienter på oral vitamin K antikoagulasjonsbehandling. Prøvematerialet er ferskt kapillært fullblod. Instrumentet produseres av iLine Microsystems S.L. og ble lansert på det skandinaviske markedet i desember 2012. Denne SKUP-evalueringen ble utført i perioden mars-august 2015 på anmodning fra iLine Microsystems S.L.

### Utprøvingen

Målet med utprøvingen var å bestemme den analytiske kvaliteten og brukervennligheten til microINR, både i bruk under optimale forhold i et sykehuslaboratorium og av brukerne i primærhelsetjenesten. Resultatene ble vurdert i forhold til kvalitetsmål satt av SKUP.

### Material og metode

Ved sykehuslaboratoriet ble det analysert kapillære prøver fra 98 pasienter, og ved de to legekantorene ble det analysert kapillære prøver fra henholdsvis 40 og 48 pasienter på microINR. Resultatene fra microINR ble sammenlignet med resultatene fra en anerkjent metode for måling av PT-INR i plasma på sykehuset. For presisjon var kvalitetsmålet en  $CV \leq 5,0 \%$  og for nøyaktighet at  $\geq 95 \%$  av resultatene fra microINR skulle avvike mindre enn 20 % fra resultatene fra sammenligningsmetoden. Kvalitetsmålet for brukervennlighet var at total vurderingen måtte være tilfredsstillende, inkludert at andelen forkastede tester forårsaket av tekniske feil måtte være  $\leq 2 \%$ . microINR ble testet i stasjonær mode; dvs. at instrumentet sto i ro både ved påføring av prøvematerialet og analysering.

### Resultat

For resultat under 2,5 INR var CV like under 5,0 % ved alle utprøvningsstedene. For resultat over 2,5 INR var CV 6,0 % under optimale forhold og mellom 6,1 og 6,3 % for brukerne i primærhelsetjenesten. Under optimale forhold ble det påvist en bias på +0,06 INR for PT-INR verdier under 2,5. For resultat over 2,5 INR ble det ikke påvist en bias. Under optimale forhold var 97 % av resultatene innenfor grensen for tillatt avvik. Hos brukerne på de to legekantorene var 95 % av resultatene innenfor grensen. Brukervennligheten ble vurdert som middels tilfredsstillende. Andel forkastede tester forårsaket av tekniske feil var på 1,6 %.

### Tilleggsinformasjon.

Fullstendig rapport fra utprøvingen av microINR, SKUP/2015/109, finnes på SKUP sin nettside [www.skup.nu](http://www.skup.nu). Laboratoriekonsulentene i Noklus kan gi råd om analysering av PT-INR på legekantor. De kan også orientere om det som finnes av alternative metoder/utstyr.

## List of previous SKUP evaluations

Summaries and complete reports from the evaluations are found at [www.skup.nu](http://www.skup.nu). In addition, SKUP reports are published at [www.skup.dk](http://www.skup.dk), where they are rated according to the national Danish quality demands for near patient instruments used in primary health care. SKUP summaries are translated into Italian by Centre for Metrological Traceability in Laboratory Medicine (CIRME), and published at <http://users.unimi.it/cirme>. SKUP as an organisation has no responsibility for publications of SKUP results on these two websites.

### The 30 latest SKUP evaluations

Evaluation no.	Component	Instrument/testkit	Producer
SKUP/2015/109	PT (INR)	microINR portable coagulometer	iLine Microsystems S.L.
SKUP/2015/108	HbA1c	<i>Confidential</i>	
SKUP/2015/106*	Strep A	QuikRead go	Orion Diagnostica Oy
SKUP/2014/101	HbA1c	InnovaStar analyzer	DiaSys Diagnostic Systems GmbH
SKUP/2014/104	PT (INR)	ProTime InRhythm	ITC International Technidyne Corporation
SKUP/2014/105	Glucose <sup>1</sup>	Accu-Chek Aviva	Roche Diagnostics
SKUP/2014/103	PT (INR)	<i>Confidential</i>	
SKUP/2013/87	Glucose <sup>1</sup>	Wella Calla Light	Med Trust Handelsges.m.b.H.
SKUP/2013/100	Glucose <sup>1</sup>	Mylife Unio	Bionime Corporation
SKUP/2013/97	NT-proBNP	Cobas h 232 POC system	Roche Diagnostics GmbH
SKUP/2013/92	CRP	Eurolyser smart 700/340	Eurolyser Diagnostica GmbH
SKUP/2013/99*	Glucose	Accu-Chek Mobile	Roche Diagnostics
SKUP/2013/98*	Glucose	Accu-Chek Aviva	Roche Diagnostics
SKUP/2013/85	Glucose, $\beta$ -Ketone	Nova StatStrip	Nova Biomedical Corporation, USA
SKUP/2013/96	Hemoglobin	DiaSpect Hemoglobin T	DiaSpect Medical GmbH
SKUP/2013/68	Allergens	ImmunoCap Rapid	Phadia AB Marknadsbolag Sverige
SKUP/2012/95	Glucose <sup>1</sup>	Mendor Discreet	Mendor Oy
SKUP/2012/94	Glucose <sup>1</sup>	Contour XT	Bayer Healthcare
SKUP/2012/91	HbA1c	Quo-Test A1c	Quoient Diagnostics Ltd
SKUP/2011/93*	Glucose	Accu-Chek Performa	Roche Diagnostics
SKUP/2011/90	CRP	<i>i</i> -Chroma	BodiTech Med. Inc.
SKUP/2011/84*	PT (INR)	Simple Simon PT and MixxoCap	Zafena AB
SKUP/2011/86	Glucose <sup>1</sup>	OneTouch Verio	LifeScan, Johnson & Johnson
SKUP/2011/77	CRP	<i>Confidential</i>	
SKUP/2011/70*	CRP	smartCRP system	Eurolyser Diagnostica GmbH
SKUP/2010/83*	Glucose	<i>Confidential</i>	
SKUP/2010/78	HbA1c	In2it	Bio-Rad
SKUP/2010/80	PT (INR)	INRatio2	Alere Inc.
SKUP/2010/89*	Glucose	FreeStyle Lite	Abbott Laboratories
SKUP/2010/88*	HbA1c	<i>Confidential</i>	

\*A report code followed by an asterisk indicates that the evaluation is not complete according to SKUP guidelines, since the part performed by the intended users was not included in the protocol, or the evaluation is a follow-up of a previous evaluation, or the evaluation is a special request from the supplier.

<sup>1</sup>Including a user-evaluation among diabetes patients.



iLine Microsystems S.L.  
 Paseo Miralelegi 50  
 20005 DUNOSTA-SAN SEBASTIÁN  
 Guipúzcoa - España  
 C.P. 4920056751  
 www.ilinemicrosystems.com  
 info@ilinemicrosystems.com  
 Tel: +34 943 005 051  
 Fax: +34 943 008 737

### SKUP Norway



November 11, 2015

### Comment to the SKUP Evaluation Report on the microINR system

Dear Ladies and Gentlemen,

Thank you for taking the time to evaluate the microINR system from iLine Microsystems.

We are very pleased with your findings. The accuracy goal set in the protocol was not only fulfilled but shows an outstanding accuracy with the Owren method.

The overall imprecision for INRs higher than 2.5 was slightly higher than the 5% goal set by SKUP.

Fully unexpected was to find that the microINR system has been intermediate rated in user-friendliness where abundant independent reports rate the microINR system above its competitors.

iLine Microsystems is pleased with the results of this study and we conclude that the microINR system is safe and reliable for managing patients under oral VKA therapy. Needless to say these findings along with other market feedback will contribute to our continuous efforts in the evolution of the microINR system.

Best regards

iLine Microsystems S.L.

Dr. Iñaki Sádaba  
 CEO

Elisabete Lasuen, MBA  
 Head of Sales and Marketing

friendly technology