

Summary of an evaluation provided by SKUP | cobas b 101 for measurement of HbA1c

Manufacturer Roche Diagnostics GmbH
Supplier Roche Diagnostics Norge AS in Norway* Roche Diagnostics A/S in Denmark*
Launched in Scandinavia April 2013



Aim

The aim of the evaluation was to assess the analytical quality and user-friendliness of **cobas b 101 HbA1c**, when used under real-life conditions by intended users in primary health care centres (PHCCs). Assessment under optimal and real-life conditions was performed under a former evaluation (SKUP/2020/117). The quality goal for repeatability was fulfilled under optimal conditions (CV 1,4-3,4%), while the quality goal for accuracy was not (83%).

Evaluated parameters	Quality goals	Conclusions and results
<i>Repeatability</i>	CV \leq 3,0 %	Fulfilled under real-life conditions (CV 1,4-2,6 %)
<i>Accuracy</i>	\geq 95 % of the results should be within \pm 3,0 mmol/mol from the results of the comparison method at HbA1c concentration $<$ 35,3 mmol/mol and within \pm 8,5 % at HbA1c concentration \geq 35,3 mmol/mol	Not fulfilled under real-life conditions (83 %)
<i>User-friendliness</i>	A total rating of "Satisfactory"	Not fulfilled (Fulfilled under real-life conditions in SKUP/2020/117)

Background

Measurement system In vitro diagnostic device for measurement of C-reactive protein (CRP), Haemoglobin A1c (HbA1c) and a Lipid Panel
Intended users Health care professionals
Sample material Capillary whole blood, or venous ethylenediaminetetraacetic acid (EDTA) or lithium-heparinised venous whole blood. Capillary whole blood was evaluated.

Material and methods

Participants 106 patients in PHCC's
Comparison method A high performance liquid chromatography (HPLC) method implemented on D-100 HbA1c System (Bio-Rad Laboratories, Inc.)
Analytical procedure The PHCC's received a demonstration of **cobas b 101 HbA1c** by Roche Diagnostics Norway.
Analysis of fresh capillary whole blood samples from each participant on **cobas b 101 HbA1c**. The measurements were performed in duplicate, i.e. two separate finger sticks. Three lots of test discs were used.
Analysis of venous samples (K₂-EDTA sample tubes) from the same individuals were measured in duplicate on the comparison method.
The evaluation was carried out from June to December 2021.
User-friendliness Assessed using a questionnaire with three given ratings; satisfactory, intermediate and unsatisfactory

Additional findings

Bias \approx +2 mmol/mol between **cobas b 101 HbA1c** and the comparison method
Technical errors None

Roche Diagnostics has accepted the report without further comments

* Requesting company